

Amended/Modified Technical Specifications

Sub.:- Submission of offers in double Bid Format as per Amended/Modified/Revised Technical Specifications.

Ref.:- Enquiry No. IICB/PUR/504/578/11/2017-18 dated 19.04.2017 for the procurement of "Bench-top, automated, digital DNA/RNA analyzing platform for NGS sample preparation"

All concerned are hereby informed that the Technical Specifications of CSIR-IICB as mentioned in Chapter No. II pertaining to Enquiry No. IICB/PUR/504/578/11/2017-18 dated 19.04.2017 is hereby amended / modified to read as mentioned below:-

Amended/Modified/Revised Technical Specifications**Specifications for "Benchtop, automated, digital DNA/RNA analyzing platform for NGS sample preparation"**

1. The system should be an automated analyzer based on microfluidics or capillary, capable to run chip or capillary electrophoresis as a part of quality control of DNA, and RNA.
2. The system must offer automated analysis of up to 1 to multiple samples in less than 30 minutes.
3. The system must be able to offer dedicated applications for NGS Workflow & Small RNA. The kits for same should be available through the same vendor.
4. The system must be able to analyze broad range of cell parameters such as Antibody staining, apoptosis detection by Annexin V binding as well as caspase-3 detection.
5. The system software must have flexibility to compare samples across multiple chips or capillary.
6. System shall have RNA quality check with RIN, offering total RNA, mRNA and Small RNA's data including RNA Integrity quality check.
7. The data of biochemical analysis should be in digital form for convenient analysis, archiving and storage.
8. Color coded result flagging tool in the software shall be available
9. System shall come with various data-display options as gel view, electropherograms and tables
10. The system should come up with a warranty of minimum 1 year.
11. System may be offered with an optional full coverage for parts and labor, reagent replacement, technical support, on-site response and advanced training

- All concerned are also informed that due date of submission of offer/quotations in Two-Bid/Double-Bid Format and opening of technical bids is extended as per details mentioned below:-

Last date and time of submission of bids/quotations by the prospective bidders in Double Bid Format	19.05.2017 up to at 1.00 PM (IST)
Date and Time of opening of Technical Bids	19.05.2017 at 3.30 PM (IST) onwards
Venue of opening of Technical Bids	Purchase Section, CSIR-IICB, Kolkata
Tenders to be submitted at	In the Tender Box kept at Stores & Purchase Section, CSIR-IICB, Jadavpur, Kolkata-32.

• All concerned may kindly also note the followings:-

- (i) Vendors may withdraw their earlier quote if they have already submitted to CSIR-IICB and want to submit the fresh quote as per revised technical specifications of CSIR-IICB. If they do not wish to submit the fresh quote (if already submitted) then their original quote will be considered for technical evaluation. Technical evaluation will be done as per this amended/modified/revised technical specifications. If two offers of the same vendor is received from the Tender Box while opening of the technical bids then the quote of the later date will be considered for evaluation. All the other terms & conditions of tender enquiry will remain unaltered.
- (ii) The vendors who have already submitted their quote with EMD and want to submit the fresh quote then they may submit the fresh quote in Two-Bid Format without EMD alongwith Technical Bid (Part-I) subject to submission of the EMD within 48 hours from the date and time of opening of technical bids failing which their technical offer will not be considered for technical evaluation and offer will be rejected.

Sunil
09.05.2017
Stores & Purchase Officer